

REPORT

TEST FACILITY

NAMSA
6750 Wales Road
Northwood, OH 43619
419.666.9455

SPONSOR

Svetlana Contrada
Berry Plastics
251 Circle Drive North
Piscataway, NJ 08854

CONFIDENTIAL

STUDY TITLE

ISO Skin Irritation Study in Rabbits

TEST ARTICLE NAME

M123 Pressure Sensitive Adhesive

TEST ARTICLE IDENTIFICATION

1064931

NAMSA

TABLE OF CONTENTS

Page

Summary	3
1. Introduction	4
2. Identification of Test and Control Articles	4
3. Test System	4
4. Animal Management	5
5. Method	6
6. Evaluation and Statistical Analysis	6
7. Results	7
8. Conclusion	7
9. Records	7
10. References	8
Appendix 1 - Classification System For Skin Reaction	9
Appendix 2 - Dermal Observations	10

Summary

The test article, M123 Pressure Sensitive Adhesive, was evaluated for primary skin irritation in rabbits. This study was conducted in accordance with the guidelines of ISO 10993-10, Biological evaluation of medical devices - Part 10: Tests for irritation and skin sensitization. Two 25 mm x 25 mm sections of the test article and control article were topically applied to the skin of each of three rabbits and left in place for a minimum of 23 hours and a maximum of 24 hours. The sites were graded for erythema and edema at 1, 24, 48 and 72 hours after removal of the single sample application.

There was no erythema and no edema observed on the skin of the animals treated with the test article. The Primary Irritation Index for the test article was calculated to be 0.0. The response of the test article was categorized as negligible.

Supervisory Personnel: Lisa A. Severhof, BA
Manager, In vivo Biocompatibility

Approved by:


Arizona E. Carter, BS, ALAT
Technical Reviewer

05-23-16
Date

Authorization for duplication of this report, except in whole, is reserved pending NAMSA's written approval.

1. Introduction

1.1 Purpose

The purpose of this study was to evaluate the test article for the potential to cause skin irritation in the rabbit.

1.2 Testing Guidelines

This study was conducted based on the International Organization for Standardization 10993-10, Biological evaluation of medical devices, Part 10: Tests for irritation and skin sensitization.

This test was performed under an ISO 13485 certified Quality System, with the test method accredited to the ISO 17025 Standard.

1.3 Dates

Test Article Received: April 14, 2016

Treatment Started: May 16, 2016

Observations Concluded: May 20, 2016

2. Identification of Test and Control Articles

The test article provided by the sponsor was identified and handled as described below:

Table 1: Test Article

Name:	M123 Pressure Sensitive Adhesive
Identification:	1064931
Physical Description of the Test Article:	M123 adhesive on supporting white paper liners. Remove both white paper release liners before the test. Test adhesive only.
Storage Conditions:	Ambient Temperature

Table 2: Control Article

Name:	Four-ply gauze, supplied by the test facility, was cut into 25 mm x 25 mm sections.
Strength, Purity, Composition or Other Characteristics:	Purity: FDA Quality System Requirements (QSR) as stipulated in 21 CFR Part 820; Composition: 20% rayon, 80% polyester blend

3. Test System

3.1 Test System

Species: Rabbit (*Oryctolagus cuniculus*)
Breed: New Zealand White
Source: Robinson Services, Inc.
Sex: Male
Body Weight: 2.7 kg at selection
Age: Young adult
Acclimation Period: Minimum 5 days
Number of Animals: Three
Identification Method: Ear tag

3.2 Justification of Test System

The rabbit (animal) is specified as an appropriate animal model for evaluating potential skin irritants by the current ISO testing standards. The rabbit is widely used for this purpose and relative ranking of irritant scores can be determined.

4. Animal Management

4.1 Husbandry, Housing and Environment

Conditions conformed to NAMSA Standard Operating Procedures that are based on the "*Guide for the Care and Use of Laboratory Animals*." Animals were individually housed in stainless steel or plastic suspended cages identified by a card indicating the lab number, animal number, test code, sex, and date dosed.

The animal housing room temperature and relative humidity were monitored daily. The temperature for the room was set to 61-72°F and the relative humidity was set to 30-70%. There were no significant temperature or relative humidity excursions that adversely affected the health of the animals.

The light cycle was controlled using an automatic timer (12 hours light, 12 hours dark).

4.2 Food, Water and Contaminants

A commercially available rabbit feed, Laboratory Rabbit Diet – 5326, was provided daily. Potable water was provided *ad libitum* through species appropriate water containers or delivered through an automatic watering system.

No contaminants present in the feed and water impacted the results of this study.

4.3 Accreditation

NAMSA is an AAALAC International accredited facility and is registered with the United States Department of Agriculture. Additionally, NAMSA maintains an approved Animal Welfare Assurance on file with the National Institutes of Health, Office for Laboratory Animal Welfare.

4.4 Personnel

Associates involved were appropriately qualified and trained.

4.5 Veterinary Care

Standard veterinary medical care was provided in this study.

4.6 IACUC

This procedure has been approved by the NAMSA Institutional Animal Care and Use Committee (IACUC), and is reviewed at least annually by the same committee.

4.7 Selection

Only healthy, previously unused, animals free from irritation or other dermatological lesions that could interfere with the test were selected.

5. Method

5.1 Test Article Preparation

The test article was an M123 adhesive on supporting white paper liners. Two 25 mm x 25 mm sections of the test article were prepared. Prior to testing the release liners were removed and either side was used for testing. The test article was backed with 4-ply gauze and applied to the skin of the rabbit.

5.2 Test Procedure

The animals were weighed and the fur on the back of each animal was clipped with an electric clipper 4 to 24 hours prior to treatment. On the day of treatment, four sites, two on each side of the back and positioned cranially and caudally, were designated on each animal. The sites were free of blemishes that could interfere with the interpretation of results.

The test article was applied to one cranial site and one caudal site (two sites per animal) by introduction under a 4 ply gauze layer to an area of skin approximately 25 mm x 25 mm square. The patches were covered with a nonreactive tape. The control was similarly applied to the opposite cranial and caudal sites. The trunk of each animal was wrapped with an elastic binder to maintain the test patches in position. Animals were returned to their cages after treatment.

After a minimum of 23 hours and a maximum of 24 hours exposure, the binders, tape, and patches were removed. The perimeter of each site was marked with nontoxic ink. The sites were gently wiped with a gauze sponge dampened with deionized water in an attempt to remove any remaining residue.

5.2.1 Laboratory Observations

1. Animals were observed daily for general health.
2. Body weights were recorded for each animal at pretreatment.
3. Dermal observations for erythema and edema were recorded at 1, 24, 48 and 72 hours after patch removal in accordance with the criteria in Appendix 1.

All times and temperatures reported herein are approximate and are within ranges established by the external standards described in the References section of this report and/or NAMSA standard operating procedures.

6. Evaluation and Statistical Analysis

The Primary Irritation Index of the test was calculated following test completion for each animal. The erythema and edema scores obtained at the 24, 48 and 72 hour intervals were added together and divided by the total number of observations. The 1 hour score was not included in the calculation. This calculation was conducted separately for the test and control article for each animal. The score for the control was subtracted from the score for the test article to obtain the Primary Irritation Score. The Primary Irritation Score for each animal was added together to determine the Combined Primary Irritation Score. The Combined Primary Irritation Score was divided by the number of animals to obtain the Primary Irritation Index. The Primary Irritation Index was characterized based on the definitions outlined in Appendix 1.

7. Results

7.1 Clinical Observations

All animals were clinically normal throughout the study.

7.2 Dermal Observations

Individual results of dermal scoring are presented in Appendix 2. No irritation was observed on the skin of the animals. No erythema and no edema reactions resulted in an overall evaluation of negligible irritation to the skin of the animals treated with the test article. The Primary Irritation Index of the test article was calculated to be 0.0. The irritation calculations are shown below:

Table 3: Irritation Calculations

Animal Number	Test Score Average	-	Control Score Average	Individual Primary Irritation Score	Combined Primary Irritation Score (CPIS)	Primary Irritation Index (CPIS ÷ 3)	Response Category
11810	0.0	-	0.0	0.0	0.0	0.0	Negligible
11811	0.0	-	0.0	0.0			
11825	0.0	-	0.0	0.0			

8. Conclusion

There was no erythema and no edema observed on the skin of the animals treated with the test article. The Primary Irritation Index for the test article was calculated to be 0.0. The response of the test article was categorized as negligible.

Results and conclusions apply only to the test article tested. Any extrapolation of these data to other articles is the sponsor's responsibility.

9. Records

All raw data pertaining to this study and a copy of the final report are retained in designated NAMSA archive files in accordance with NAMSA SOPs.

10. References

Code of Federal Regulations (CFR), Title 9, Parts 1-4, Animal Welfare Act.

Code of Federal Regulations (CFR), Title 16, Part 1500, Federal Hazardous Substances Act (FHSA) Regulations (2008).

International Organization for Standardization (ISO) 10993-1, Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process (2009/Technical Corrigendum 1 2010).

International Organization for Standardization (ISO) 10993-2, Biological evaluation of medical devices - Part 2: Animal welfare requirements (2006).

International Organization for Standardization (ISO) 10993-10, Biological evaluation of medical devices - Part 10: Tests for irritation and skin sensitization (2010).

International Organization for Standardization (ISO) 10993-12, Biological evaluation of medical devices - Part 12: Sample preparation and reference materials (2012).

National Research Council, *Guide for the Care and Use of Laboratory Animals*, Washington, DC: National Academy Press, 2011.

Office of Laboratory Animal Welfare (OLAW), Public Health Service Policy on Humane Care and Use of Laboratory Animals.

Appendix 1 - Classification System For Skin Reaction

Reaction	Numerical Grading
Erythema and Eschar Formation	
No erythema	0
Very slight erythema (barely perceptible)	1
Well-defined erythema	2
Moderate erythema	3
Severe erythema (beet redness) to eschar formation preventing grading of erythema	4
Edema Formation	
No edema	0
Very slight edema (barely perceptible)	1
Well-defined edema (edges of area well-defined by definite raising)	2
Moderate edema (raised approximately 1 mm)	3
Severe edema (raised more than 1 mm and extending beyond exposure area)	4
Total possible score for irritation	8

NOTE: Other adverse changes at the skin sites shall be recorded and reported

Irritation Response Categories in the Rabbit

Response Category	Mean Score
Negligible	0.0 to 0.4
Slight	0.5 to 1.9
Moderate	2.0 to 4.9
Severe	5.0 to 8.0

Appendix 2 - Dermal Observations

Animal Number/ Sex	Weight (kg)	Group	Observation	Interval (hours)							
				1		24		48		72	
				Left	Right	Left	Right	Left	Right	Left	Right
11810 Male	2.7	Test	Erythema	0	0	0	0	0	0	0	0
			Edema	0	0	0	0	0	0	0	0
		Control	Erythema	0	0	0	0	0	0	0	0
			Edema	0	0	0	0	0	0	0	0
11811 Male	2.7	Test	Erythema	0	0	0	0	0	0	0	0
			Edema	0	0	0	0	0	0	0	0
		Control	Erythema	0	0	0	0	0	0	0	0
			Edema	0	0	0	0	0	0	0	0
11825 Male	2.7	Test	Erythema	0	0	0	0	0	0	0	0
			Edema	0	0	0	0	0	0	0	0
		Control	Erythema	0	0	0	0	0	0	0	0
			Edema	0	0	0	0	0	0	0	0